

Bhagat Phool Singh Govt. Medical College, Khanpur Kalan, Sonipat

Email: bpsgmc.purchase@gmail.com

Invitation for Sealed quotation for Table top

Centrifuge. (To be substitute with quotation called earlier vide letter no. pur/25/8380 dated-12-08-2025)

Inquiry No: Purchase/25/

Inquiry Issue Date:

Last Date of Submission: 10.09.2025 at 11:00 AM

Sealed/e-Quotations are hereby invited by the undersigned on behalf of the Director, BPSGMC for Women, Khanpur Kalan, Sonapat for repairing as per **Annexure-1** for the Institute as per terms & conditions mentioned below.

The sealed/e-quotations received reach in the office of the undersigned on or before 10.09.2025 at 11:00 AM. The Envelope containing the quotation would please be sealed and super scribed as under:-

1. Terms & Conditions:

- a. The quotations received after this deadline & unsealed projected shall not be entertained under any circumstances whatsoever. In case of any delay this Institute will not be responsible.
- b. Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be submitted in the office of undersigned before deadline of submitting the quotation.
- c. Rates must be quoted in **Indian Rupees** and as per the format specified taxes extra if any must be written separately.
- d. Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- e. No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- f. The rates quoted must be valid as per department and State Govt. instruction issued from time to time, from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- g. Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- h. Any conditional quotation shall be rejected summarily.
- i. **Payment Terms:** Payment will be only after satisfactorily completion of work /commissioning of material and after inspection by Inspection Committee.
- j. BPSGMC Khanpur Kalan, Sonapat reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the BPSGMC Khanpur Kalan, Sonapat will be final in this regard.
- k. BPSGMC Khanpur Kalan, Sonapat reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the BPSGMC, Sonapat will be final in this regard.

2 Special Terms & Conditions:

- a) Bidder must quote the product as per specification provided in Annexure 1.
- b) Inspection committee will check the work thoroughly, if somehow inspection committee found any discrepancy and is not satisfied with the final product (final product should be same as reviewed ones at the institution) then BPSGMC Khanpur Kalan, Sonipat has the right to reject the supply and cancel the order, and no claim for payment in this regard will be entertained


Accounts Officer

Encl.: Annexure 1 (Demand)

Annexure 2 (Format of price bid)

Annexure-1

Sr. No	Name of the Item	Terms and conditions
1	Table top Centrifuge in Department of blood Bank	1. Provide at least five years warranty and five years CMC items purchased (copy attached <i>for specifications</i>)

ANNEXURE "2"**[On the letterhead of firm]****PRICE BIDFORM**

To,
Director Officer,
BPSGMC,
Khanpur Kalan, Sonipat

1. I/We Submitted the quotation for Enquiry No. "QUOTATIONFOR SUPPLY..... AGAINST THE INQUIRY NO: Purchase/25/.... due on dated.....at BPSGMC, Khapur Kalan, Sonipat.
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

Sr. No	Particular	Quantity	Quoted Make	Price/Unit Exclusive of Tax (INR)	GST/CST/ST
1.					
2.					

Date _____

Place _____

(Signature _____ of _____ Authorized
Person) _____

(Name) _____

Phone No: _____

BPS, GMC FOR WOMEN, KHANPUR W
Department of Immunohematology / Blood Transfusion

Laboratory Table Top Centrifuge

TECHNICAL SPECIFICATIONS for Laboratory Centrifuge (24 TUBES)
Laboratory Centrifuges should be suitable for routine sample analysis in blood bank laboratories
Should have stable speed output even under unstable voltage conditions Low sample temperature rise

Should have Brushless Induction motor with frequency drive
Should have Microprocessor controller with touch button LED display

Should have Smooth & soft start
Should have Inverter fault detection with auto shutdown

Should have Dynamic brake for quick deceleration
Should have Last set parameters recall (Useful for repetitive analysis)

Should have Stainless Steel centrifuge Chamber, easy to clean
Should have Safety lid interlock to prevent lid opening during centrifugation

Should have imbalance detection & centrifugation stop with display of error
Should not require puller for changing the rotor head

Should have Calibration window on the lid
Should have Motor overload protection

Should have Gas hinge to prevent door falling
Should have Emergency lid lock release

Should have Feather touch setting keys
Swing out rotor with capacity of minimum 24 blood tubes

Should have Alphanumeric LED display of speed & Time
Should have Max. Speed upto 4500 RPM

Should have Max. RCF upto 3300 g
Should have Digital timer range 0-99 min
Should be USFDA / European CE / BIS certified.

Five years warranty and Five years CMC
Give reference at least two institution installation

7
mail. CF.